

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Disinfectant											
	Cidex	Cidex OPA	Gigasept AF	Gigasept FF	Virkon	Cidex Plus	Wavicide-01	AIDAL PLUS	Cetylcode-G	Hibitane	Metricide 14	Sporicidin
Manufacturer	Johnson and Johnson	Johnson and Johnson	Schülke&Mayr UK Ltd.	Schülke&Mayr UK Ltd.	Dupont (Antec)	Johnson and Johnson	Medical Chemical Corp.	Whiteley Medical		COOPERS	Metrex	CONTEC
Level	HLD	HLD	-	HLD	HLD	HLD	HLD	HLD		-	HLD	HLD
Active Ingredients	Glutaraldehyde	Orthophthalaldehyde	Succinic dialdehyde	Succinic dialdehyde	Potassium	Glutaraldehyde	Glutaraldehyde	Glutaraldehyde	Glutaraldehyde	Clorhexidine gluconate	Glutaraldehyde	Phenol
FDA Approval	Y	Y	Y	Y	Y	Y	Y	Y	Y	N	Y	Y
TYPE	Liquid	Liquid	Liquid	Liquid	Powder	Liquid	Liquid	Liquid	Liquid	Liquid	Liquid	Liquid
Transducer	C1-6	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	C1-6T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	C1-6i	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	C1-6CT	Δ	O	Δ	X	Δ	Δ	O	Δ	Δ	Δ	O
	SC1-6	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	C5-8	O	O	O	O	Δ	O	O	Δ	Δ	O	Δ
	SC1-4HS	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	SC1-4H	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	SC1-6H	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-8	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-8i	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12i	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12H	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12HWD	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L3-12X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L8-17	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L8-17i	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	L8-17X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	MP1-5X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	SP1-5	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	SP1-5i	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	SP3-8	Δ	O	Δ	O	O	O	O	Δ	Δ	Δ	X
	SP1-5X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	P1-5CT	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	VC1-6	Δ	O	Δ	O	O	O	Δ	O	Δ	Δ	Δ
	VC1-6T	Δ	O	Δ	O	O	O	Δ	O	Δ	Δ	Δ
	SVC1-6	Δ	O	Δ	O	O	O	Δ	O	Δ	Δ	Δ
	SVC1-6H	Δ	O	Δ	O	O	O	Δ	O	Δ	Δ	Δ
	E3-10	Δ	O	Δ	Δ	O	O	Δ	Δ	Δ	Δ	Δ
	EN3-10	O	O	Δ	Δ	Δ	O	O	Δ	Δ	Δ	Δ
	EV3-10	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EV3-10T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EC3-10	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EC3-10T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	E3-10H	Δ	O	Δ	Δ	O	O	Δ	Δ	Δ	Δ	Δ
	EC3-10H	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EV3-10H	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	IO3-12	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	CW2.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	IO8-17	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EV3-10X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
	EC3-10X	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X
VE3-10H	Δ	O	Δ	O	O	O	Δ	O	Δ	Δ	Δ	
C5-8N	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
C5-8NT	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
SP1-5T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
SP3-8T	Δ	O	Δ	O	O	O	O	Δ	Δ	Δ	X	
IO8-17T	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
minisono L3-12	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
minisono C1-6	Δ	O	Δ	X	O	O	O	Δ	Δ	Δ	X	
Accessory	US-Guider	Δ	O	Δ	X	X	O	O	Δ	Δ	Δ	O

※ CAUTION ※

- Do not wipe the cable, strain relief, and connector of the transducer with the disinfectant, as it can damage and/or discolor those parts of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless they are sterilizable.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Disinfectant											
	Omnicide -FG2	Nuclean	Ster-Bac Blu	Sani-Cloth HB	PeraSafe	Super Sani-Cloth	Cleanisept wipes	CLEANISEPT® WIPES forte	Protex	Tristel Duo	SteriMax® Sporicide Wipes	SteriMax® Wipes
Manufacturer	Interhatch	Ecolab	Ecolab	NICE PAK	VWR	NICE PAK	Dr. Schumacher GmbH	Dr. Schumacher GmbH	PARKER	Tristel solutions	Aseptix	Aseptix
Level	-	-	-	LLD	HLD	-	LLD	LLD	LLD	HLD	-	-
Active Ingredients	Glutaraldehyde	Glutaraldehyde	Quaternary Ammonium	Quaternary Ammonium	Peracetic	Alcohol	Quaternary Ammonium	Quaternary Ammonium	Quaternary Ammonium	Chlorine Dioxide	Hydrogen peroxide	Hydrogen peroxide
FDA Approval	Y	N	N	N	N	N	N	N	Y	Y	N	N
TYPE	Liquid	Liquid	Liquid	Wipes	Liquid	Wipes	Wipes	Wipes	Spray	Spray	Wipes	Wipes
Transducer	C1-6	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	C1-6T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	C1-6i	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	C1-6CT	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	O	O
	SC1-6	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	C5-8	Δ	Δ	O	O	O	Δ	Δ	Δ	Δ	Δ	Δ
	SC1-4HS	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	SC1-4H	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	SC1-6H	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-8	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-8i	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12i	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12H	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12HWD	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L3-12X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L8-17	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L8-17i	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	L8-17X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	MP1-5X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	SP1-5	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	SP1-5i	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	SP3-8	Δ	Δ	Δ	O	Δ	O	Δ	Δ	Δ	Δ	X
	SP1-5X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	P1-5CT	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	VC1-6	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	O	Δ
	VC1-6T	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	O	Δ
	SVC1-6	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	O	Δ
	SVC1-6H	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	O	Δ
	E3-10	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	Δ	Δ
	EN3-10	O	O	Δ	Δ	Δ	O	Δ	Δ	Δ	Δ	Δ
	EV3-10	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EV3-10T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EC3-10	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EC3-10T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	E3-10H	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	Δ	Δ
	EC3-10H	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EV3-10H	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	IO3-12	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	CW2.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	IO8-17	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EV3-10X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	EC3-10X	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
	VE3-10H	Δ	Δ	Δ	Δ	Δ	O	O	O	Δ	O	Δ
	C5-8N	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X
C5-8NT	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X	
SP1-5T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X	
SP3-8T	Δ	Δ	Δ	O	Δ	O	Δ	Δ	Δ	Δ	X	
IO8-17T	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X	
minisono L3-12	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X	
minisono C1-6	Δ	Δ	Δ	Δ	Δ	O	O	O	O	O	X	
Accessory	US-Guider	Δ	Δ	Δ	Δ	Δ	O	O	O	X	X	X

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Disinfectant											
	SterilMax Sporicide	UltraSan Broad Spectrum	CaviCide	CaviWipes	Metricide OPA Plus	Metricide	Meliseptol Wipes sensitive	Meliseptol Wipes ultra	Revital-OX Resert	Meliseptol Foam pure	Sporox II	Protex wipes
Manufacturer	Aseptix	Aseptix	Metrex	Metrex	Metrex	Metrex	B. Braun	B. Braun	STERIS	B. Braun	SultanHealthcare	PARKER
Level	-	-	-	LLD,ILD	HLD	HLD	-	-	HLD	-	HLD	-
Active Ingredients	Hydrogen peroxide	Hydrogen peroxide	Isopropanol	Isopropanol	Orthophthalaldehyde	Glutaraldehyde	Propan-1-ol	Quaternary Ammonium	Hydrogen Peroxide	Propan-1-ol	Hydrogen Peroxide	Ammonium chloride
FDA Approval	N	N	Y	Y	Y	Y	N	N	Y	N	Y	Y
TYPE	Spray	Spray	Spray	Wipes	Liquid	Liquid	Wipes	Wipes	Liquid	Liquid	Liquid	Wipes
Transducer	C1-6	X	X	X	O	O	X	O	O	O	O	O
	C1-6T	X	X	X	O	O	X	O	O	O	O	O
	C1-6i	X	X	X	O	O	X	O	O	O	O	O
	C1-6CT	Δ	Δ	O	O	Δ	Δ	Δ	O	O	X	O
	SC1-6	X	X	X	O	O	X	O	O	O	O	O
	C5-8	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	SC1-4HS	X	X	X	O	O	X	O	O	O	O	O
	SC1-4H	X	X	X	O	O	X	O	O	O	O	O
	SC1-6H	X	X	X	O	O	X	O	O	O	O	O
	L3-8	X	X	X	O	O	X	O	O	O	O	O
	L3-8i	X	X	X	O	O	X	O	O	O	O	O
	L3-12	X	X	X	O	O	X	O	O	O	O	O
	L3-12T	X	X	X	O	O	X	O	O	O	O	O
	L3-12i	X	X	X	O	O	X	O	O	O	O	O
	L3-12H	X	X	X	O	O	X	O	O	O	O	O
	L3-12HWD	X	X	X	O	O	X	O	O	O	O	O
	L3-12X	X	X	X	O	O	X	O	O	O	O	O
	L8-17	X	X	X	O	O	X	O	O	O	O	O
	L8-17i	X	X	X	O	O	X	O	O	O	O	O
	L8-17X	X	X	X	O	O	X	O	O	O	O	O
	MP1-5X	X	X	X	O	O	X	O	O	O	O	O
	SP1-5	X	X	X	O	O	X	O	O	O	O	O
	SP1-5i	X	X	X	O	O	X	O	O	O	O	O
	SP3-8	X	X	X	Δ	Δ	X	Δ	Δ	Δ	Δ	Δ
	SP1-5X	X	X	X	O	O	X	O	O	O	O	O
	P1-5CT	X	X	X	O	O	X	O	O	O	O	O
	VC1-6	Δ	Δ	Δ	Δ	O	Δ	O	O	O	O	O
	VC1-6T	Δ	Δ	Δ	Δ	O	Δ	O	O	O	O	O
	SVC1-6	Δ	Δ	Δ	Δ	O	Δ	O	O	O	O	O
	SVC1-6H	Δ	Δ	Δ	Δ	O	Δ	O	O	O	O	O
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	EN3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	X
	EV3-10	X	X	X	O	O	X	O	O	O	O	O
	EV3-10T	X	X	X	O	O	X	O	O	O	O	O
	EC3-10	X	X	X	O	O	X	O	O	O	O	O
	EC3-10T	X	X	X	O	O	X	O	O	O	O	O
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	EC3-10H	X	X	X	O	O	X	O	O	O	O	O
	EV3-10H	X	X	X	O	O	X	O	O	O	O	O
	IO3-12	X	X	X	O	O	X	O	O	O	O	O
	CW2.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	IO8-17	X	X	X	O	O	X	O	O	O	O	O
	EV3-10X	X	X	X	O	O	X	O	O	O	O	O
	EC3-10X	X	X	X	O	O	X	O	O	O	O	O
	VE3-10H	Δ	Δ	Δ	Δ	O	Δ	O	O	O	O	O
	C5-8N	X	X	X	O	O	X	O	O	O	O	O
C5-8NT	X	X	X	O	O	X	O	O	O	O	O	
SP1-5T	X	X	X	O	O	X	O	O	O	O	O	
SP3-8T	X	X	X	Δ	Δ	X	Δ	Δ	Δ	Δ	Δ	
IO8-17T	X	X	X	O	O	X	O	O	O	O	O	
minisono L3-12	X	X	X	O	O	X	O	O	O	O	O	
minisono C1-6	X	X	X	O	O	X	O	O	O	O	O	
Accessory	US-Guider	X	X	X	O	O	X	Δ	Δ	Δ	Δ	Δ

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Disinfectant							Pre-Cleaner					
	Sani-cloth AF3	Acrylan	Descogen Liquid	Cutasept F	Cutasept G	Mikrobac Virucidal Tissues	SONO Ultrasound and Disinfecting Wipe	T-spray	T-spray II	Enzol	Theracide Plus(S5N)	Transeptic	
Manufacturer	NICE PAK	ANTISEPTICA	ANTISEPTICA	BODE Chemie GmbH	BODE Chemie GmbH	BODE Chemie GmbH	ADVANCED ULTRASOUND SOLUTIONS INC.	Pharmaceutical Innovations, INC	Pharmaceutical Innovations, INC	Johnson and Johnson	Burnishine Products	PARKER	
Level	LLD,ILD	-	-	-	-	-	-	-	-	-	-	-	
Active Ingredients	Quaternary Ammonium	Quaternary ammonium compounds	Potassium peroxymonosulfate	Propan-2-ol	Propan-2-ol	ammonium chloride	ammonium chloride	Alkyl dimethyl benzyl ammonium chloride	Alkyl dimethyl benzyl ammonium chloride	Subtilisins(protolytic enzymes)	Isopropyl alcohol	Isopropyl alcohol	
FDA Approval	Y	N	N	N	N	N	N	Y	Y	Y	N	Y	
TYPE	Wipes	Liquid	Liquid	Liquid	Liquid	Wipes	Wipes	Spray	Spray	Liquid	Liquid	Liquid	
Transducer	C1-6	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	C1-6T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	C1-6i	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	C1-6CT	O	Δ	Δ	Δ	X	O	O	Δ	Δ	Δ	Δ	Δ
	SC1-6	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	C5-8	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	O	Δ	Δ	Δ
	SC1-4HS	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	SC1-4H	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	SC1-6H	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-8	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-8i	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12i	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12H	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12HWD	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L3-12X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L8-17	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L8-17i	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	L8-17X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	MP1-5X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	SP1-5	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	SP1-5i	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	SP3-8	Δ	X	Δ	X	X	Δ	X	Δ	Δ	O	Δ	Δ
	SP1-5X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	P1-5CT	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	VC1-6	O	O	O	O	Δ	O	O	Δ	O	O	O	O
	VC1-6T	O	O	O	O	Δ	O	O	Δ	O	O	O	O
	SVC1-6	O	O	O	O	Δ	O	O	Δ	O	O	O	O
	SVC1-6H	O	O	O	O	Δ	O	O	Δ	O	O	O	O
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	O
	EN3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	Δ
	EV3-10	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EV3-10T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EC3-10	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EC3-10T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	Δ
	EC3-10H	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EV3-10H	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	IO3-12	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	CW2.0	O	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	CW5.0	O	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	IO8-17	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EV3-10X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
	EC3-10X	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ
VE3-10H	O	O	O	O	Δ	O	O	Δ	O	O	O	O	
C5-8N	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
C5-8NT	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
SP1-5T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
SP3-8T	Δ	X	Δ	X	X	Δ	X	Δ	Δ	O	Δ	Δ	
IO8-17T	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
minisono L3-12	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
minisono C1-6	O	X	O	X	X	O	X	Δ	Δ	O	Δ	Δ	
Accessory	US-Guider	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	Δ	O	

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Pre-Cleaner							Gel				
	Metrizyme	Klenzyme	McKesson	Cidezyme	Endozyme	EmPower	Ultrasound probe cleaning wipes	Aquasonic100	ULTRASOUND GEL	Ultrasound Transmission gel	Scan	Sonogel
Manufacturer	Metrex	STERIS	McKesson	Johnson and Johnson	RUHOF Corporation	Metrex	Nanosonics	PARKER	Ultragel Hungary 2000 Ltd.	Ecosonic	PARKER	SONOGEL Vertriebs GmbH
Level	-	-	-	-	-	-	-	-	-	-	-	-
Active Ingredients	Propylene glycol	Enzymes	PCMX (Chloroxylenol)	Proteolytic Enzymes	Propanol	Propylene glycol	PHENOXYETHA NOL	-	-	-	-	-
FDA Approval	Y	Y	Y	Y	Y	Y	N	N	N	N	N	N
TYPE	Liquid	Liquid	Liquid	Liquid	Liquid	Liquid	Wipes	Gel	Gel	Gel	Gel	Gel
Transducer	C1-6	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	C1-6T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	C1-6i	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	C1-6CT	O	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	Δ	Δ
	SC1-6	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	C5-8	Δ	Δ	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	O
	SC1-4HS	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	SC1-4H	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	SC1-6H	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-8	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-8i	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12i	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12H	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12HWD	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L3-12X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L8-17	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L8-17i	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	L8-17X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	MP1-5X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	SP1-5	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	SP1-5i	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	SP3-8	O	O	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ
	SP1-5X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	P1-5CT	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	VC1-6	O	Δ	Δ	Δ	O	O	O	O	Δ	Δ	O
	VC1-6T	O	Δ	Δ	Δ	O	O	O	O	Δ	Δ	O
	SVC1-6	O	Δ	Δ	Δ	O	O	O	O	Δ	Δ	O
	SVC1-6H	O	Δ	Δ	Δ	O	O	O	O	Δ	Δ	O
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	EN3-10	O	O	O	O	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	EV3-10	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EV3-10T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EC3-10	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EC3-10T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	EC3-10H	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EV3-10H	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	IO3-12	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	CW2.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	IO8-17	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EV3-10X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	EC3-10X	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
	VE3-10H	O	Δ	Δ	Δ	O	O	O	O	Δ	Δ	O
	C5-8N	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O
C5-8NT	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O	
SP1-5T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O	
SP3-8T	O	O	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	
IO8-17T	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O	
minisono L3-12	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O	
minisono C1-6	O	Δ	Δ	Δ	Δ	O	O	O	O	Δ	O	
Accessory	US-Guider	X	Δ	Δ	Δ	Δ	Δ	O	Δ	Δ	Δ	Δ

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8

Transducer Disinfectant Material Compatibility Table

[Rev.18]



O : Approved / X : Not Approved / Δ : Not Tested

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

Product Type	Gel	Trophon® EPR	Antigermix ^(Note)	V-PRO ^(Note)
	Super sonic			
Manufacturer	Sung Heung Corporation			
Level	-			
Active Ingredients	-			
FDA Approval	N			
TYPE	Gel	Device	Device	Device
Transducer	C1-6	O	O	O
	C1-6T	O	O	O
	C1-6i	O	O	O
	C1-6CT	Δ	Δ	Δ
	SC1-6	O	O	O
	C5-8	Δ	Δ	Δ
	SC1-4HS	O	O	O
	SC1-4H	O	O	O
	SC1-6H	O	O	O
	L3-8	O	O	O
	L3-8i	O	O	O
	L3-12	O	O	O
	L3-12T	O	O	O
	L3-12i	O	O	O
	L3-12H	O	O	O
	L3-12HWD	O	O	O
	L3-12X	O	O	O
	L8-17	O	O	O
	L8-17i	O	O	O
	L8-17X	O	O	O
	MP1-5X	O	O	O
	SP1-5	O	O	O
	SP1-5i	O	O	O
	SP3-8	Δ	Δ	Δ
	SP1-5X	O	O	O
	P1-5CT	O	O	O
	VC1-6	Δ	O	X
	VC1-6T	Δ	O	X
	SVC1-6	Δ	O	X
	SVC1-6H	Δ	O	X
	E3-10	Δ	Δ	Δ
	EN3-10	Δ	Δ	Δ
	EV3-10	O	X	O
	EV3-10T	O	X	O
	EC3-10	O	O	O
	EC3-10T	O	O	O
	E3-10H	Δ	Δ	Δ
	EC3-10H	O	O	O
	EV3-10H	O	X	O
	IO3-12	O	O	O
	CW2.0	Δ	Δ	Δ
	CW5.0	Δ	Δ	Δ
	IO8-17	O	O	O
	EV3-10X	O	X	O
	EC3-10X	O	O	O
	VE3-10H	Δ	O	X
	C5-8N	O	O	O
	C5-8NT	O	O	O
	SP1-5T	O	O	O
	SP3-8T	Δ	Δ	Δ
IO8-17T	O	O	O	
minisono L3-12	O	X	O	
minisono C1-6	O	X	O	
Accessory	US-Guider	X	Δ	Δ

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2019 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E1910-V1.8